	Application No.	Applicant(s)
Notice of Allowability	10/770,483	KOBAYASHI, HIROYUKI
	Examiner	Art Unit
	Allen Wong	2621
The MAILING DATE of this communication app All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85 NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT R of the Office or upon petition by the applicant. See 37 CFR 1.31	ears on the cover sheet w. (OR REMAINS) CLOSED in or other appropriate commatter. This application is	ith the correspondence address in this application. If not included funication will be mailed in due course. THIS
1. This communication is responsive to		
2. The allowed claim(s) is/are <u>1-3</u> .		
3. ☐ Acknowledgment is made of a claim for foreign priority u a) ☐ All b) ☐ Some* c) ☐ None of the:		or (f).
1. Certified copies of the priority documents have been received.		
2. Certified copies of the priority documents have been received in Application No		
3. Copies of the certified copies of the priority documents have been received in this national stage application from the		
International Bureau (PCT Rule 17.2(a)). * Certified copies not received:		
Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. THIS THREE-MONTH PERIOD IS NOT EXTENDABLE. 4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF		
INFORMAL PATENT APPLICATION (PTO-152) which give		or declaration is deticient.
5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted.(a) including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached		
1) hereto or 2) to Paper No./Mail Date		
(b) ☐ including changes required by the attached Examiner Paper No./Mail Date	_	or in the Office action of
Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).		
6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.		
: •		
Attachment(s)		
1. Notice of References Cited (PTO-892)		nformal Patent Application
2. Notice of Draftperson's Patent Drawing Review (PTO-948)	Paper No	Summary (PTO-413), ./Mail Date
Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date	7. ∐ Examiner's	s Amendment/Comment
4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	8. 🛛 Examiner's	s Statement of Reasons for Allowance
or biological material	9. 🗌 Other	
		Allen Wong
	·	Primary Examiner Art Unit 2621

Application/Control Number: 10/770,483

Art Unit: 2621

DETAILED ACTION

Allowable Subject Matter

1. Claims 1-3 are allowed.

The following is an examiner's statement of reasons for allowance: The prior art does not specifically disclose the specific combination of limitations of claim 1: a video endoscope apparatus comprising: a light source unit that alternately emits white light covering the full range of visible region and excitation light to excite living tissue; a video endoscope that has an insertion tube, a color image sensor and a light guide for guiding the white light and the excitation light from said light source unit to a distal end of said insertion tube; an image data acquiring section that acquires normal observation image data captured by said color image sensor when said light source unit emits the white light and acquires excitation observation image data captured by said color image sensor when said light source section emits the excitation light; a first image processing section that processes said normal observation image data only acquired by said image data acquiring section; a second image processing section that extracts one color component image data as reference image data from said normal observation image data acquired by said image data acquiring section, extracts one color component image data as fluorescent image data from said excitation observation image data acquired by said image data acquiring section, calculates a difference between a brightness value of a pixel in said reference image data and a brightness value of the pixel at the same coordinate in said fluorescent image data for every pixel and generates affected part information by specifying the coordinates

Art Unit: 2621

whose differences are larger than the predetermined threshold value; an image data generating section that generates image data for displaying a special observation image by converting said normal observation image data based on said affected part information so that the values of the pixels specified by said affected part information in said normal observation image data are converted into a predetermined level corresponding to a predetermined color; and an output section that outputs said image data generated by said image data generating section.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Contact Information

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Allen Wong whose telephone number is (571) 272-7341. The examiner can normally be reached on Mondays to Thursdays from 8am-6pm Flextime.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, John W. Miller can be reached on (571) 272-7353. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR.

Art Unit: 2621

Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Allen Wong

Primary Examiner
Art Unit 2621

AW 6/19/07